

Full Length Research Paper

Development of a sensitive spectrophotometric method for determination of copper

Homayon Ahmad Panahi^{1*}, Meghdad Karimi¹, Elham Moniri², Hamidreza Soudi¹

¹Department of Chemistry, Islamic Azad University, central Tehran campus, Iran.

²Department of Chemistry, Islamic Azad University, Varamin (Pishva) campus, Iran.

Accepted 17 October, 2008

Simple and rapid spectrophotometric procedure have been established for quantitation of Cu (II). This method is based on the reaction between alizarin red S and boric acid in the solution of pH 8.5 to produce a red complex which place between organic and aqueous phases. In presence of Cu (II), the complex of alizarin red S and boric acid is dissociated and new complex between alizarin red S and metal ions formed which can be measured at the optimum wavelength. All variables were studied to optimize the reaction conditions. Regression analysis of beer's low plot showed good correlation in concentration range 0.6 - 15 $\mu\text{g.mL}^{-1}$. The molar absorptivity and detection limit are calculated. The interference effects of the foreign cation have been examined.

Key words: Extraction, copper, trace elements, water samples, Alizarin red S

INTRODUCTION

The direct determination of heavy metals at trace level in saline samples by flame atomic absorption spectrometer (FAAS) or electrothermal atomic absorption spectrometer (ETAAS) (Stalikas, 2002; Ghaedi et al., 2005; Jahromi et al., 2007; Sakai et al., 1998) has been failed due to particularly their low concentrations and matrix effects. Many authors indicated that matrix interferences are possible and to overcome them, the addition of different matrix modifiers (Ghaedi et al., 2006) as well as the use of preconcentration-separation procedures (Ghaedi et al., 2006; Stafilov, 2000; Nakamura et al., 1988; Lazaru et al., 1993; Dadfarnia et al., 2006). Preconcentration and separation techniques, such as liquid-liquid extraction, ion exchange, co-precipitation, cloud point extraction, solid phase extraction, etc. can solve these problems and lead to a higher confidence level and easy determination of the trace elements (Ghaedi et al., 2006; Stafilov, 2000; Nakamura et al., 1988; Lazaru et al., 1993; Dadfarnia et al., 2006; Rojas et al., 2007; Lahiri et al., 2007; Yaman et al., 2006; Li et al., 2007). Some of these methods suffer from inconveniences such as, lengthy separation, limitation of the volume of sample solution investigated, time

consuming, multi stage, lower enrichment factor and consumption of organic harmful solvents. These problems can be overcome by replacement of filtration or centrifugation by flotation. The major advantages of the flotation preconcentration method are the rapidity in addition to excellent recoveries of analytes. The equipments necessary for flotation preconcentration are simple and inexpensive. The small amount of a surfactant and the tiny air bubbles require to perform the proper flotation prevent serious contamination risks, which could be manifested by the high blank value. Because the physical and chemical properties of aragonite aqueous solution are similar to those of the natural waters with higher water hardness, a method of flotation is applied as the second way to eliminate the matrix effects (Aki et al., 2006; Lazaru et al., 1997; Stafilov et al., 2000). Several methods have been developed for the flotation of heavy metal ions (Caballero et al., 1990; Cundeva et al., 2000; Paneva et al., 2005; Zendelovska et al., 2001; Karimi et al., 2008; Zhao et al., 2008).

Copper determination in water and biological materials is a good tool for environmental and toxicological monitoring (Duran, 2007; Prabhakaran and Subramanian, 2003; Lemos, 2003; Guo, 2004; Tewari and Singh, 2001; Kumar, 2000; Venkatesh and Singh, 2005; Safavi et al., 2006). Generally determination of trace metal is very important in the context of environmental Protection, food

*Corresponding author. E-mail: Panahi20002000@yahoo.com.
Tel: 0098 021 88420087.




Figure 1. Structure of Alizarin red S




Figure 2. Cyclic ester formed from the reaction of 1, 2-diols with borate.

and agricultural chemistry and high purity materials. Flame atomic absorption spectrometry and spectrophotometric have been widely used for the determination of trace metal ions because of the relatively simple and inexpensive equipment required.

The application of UV-Vis absorption to the determination of metals is still popular in many laboratories. The technique provides easy determination of many metals from low to high concentration at affordable cost.

Many new organic reagents have been synthesized and various highly sensitive methods developed with molar absorptivities of 10^5 - 10^6 L. mol⁻¹.cm⁻¹ or even higher. In studies for establishing new methods, much attention paid to sensitivity. However specificity and selectivity are also very important.

In this paper UV-Vis spectrophotometric method is proposed for determination of Cu (II) in aqueous solution. The method based on the reaction Cu (II) and alizarin red S (Figure 1). The proposed method is simple, sensitive and the complex formed is stable.

EXPRIMENTAL

Materials

Chloroform, boric acid, sodium monohydrogen phosphate, sodium dihydrogen phosphate, mercury (II) chloride, silver sulfide, copper (II) sulfate pentahydrate, nickel (II) chloride hexahydrate, cobalt (II) sulfate heptahydrate, lead (II) acetate trihydrate and tin (II) chloride dehydrate were products of Merck (Darmstadt, Germany). Alizarin red S was obtained from Fluka (Buchs, Switzerland). The stock solution (1000 mg. L⁻¹) of Cu (II) was prepared by dissolving appropriate amount of Cu (SO₄)₂ .5H₂O in deionized water. 0.01 M phosphate buffer was used to adjust the pH of the solutions,

wherever suitable.

Assay procedure

3 mL alizarin red S (10 ppm), 2 mL chloroform, 1mL buffer pH 8.5 and 3 mL of boric acid 0.01 M were transferred into a test tube. The contents were shaken for approximately 1 min. The two phases were allowed to separate complex of alizarin-boric acid was formed between two aqueous and organic phases. Then aliquot phases were decanted and complex was washed several times with deionized water till obtaining colorless solution. Aliquots of solution (3 mL) of Cu²⁺ 0.6 - 15 µg.mL⁻¹ and 1 mL buffer (pH 7.5) were added into the test tube and shaken for 1 min. After separation of two phases, the absorbance of aqueous solution was measured at 240 nm against the reagent blank. The reagent blank was prepared exactly like the procedure described above, but in the absence of Cu²⁺. A calibration graph was drawn or regression equation calculated.

RESULT AND DISCUSSION

Compounds that contain hydroxylic group can react with boronate. The interaction that may occur is the formation of a cyclic ester complex, as illustrated in Figure 2. These complexes can form with 1, 2 or 1, 3-diol- containing compounds when the hydroxyl groups are oriented in the proper geometry like alizarin red S.

Due to the simple stoichiometry of the occurred reaction dealing with the formation of boric acid with alizarin red S and also high stability and insolubility of obtained precipitate in aqueous media, we were interested in developing a new simple method for the quantitative determination of Cu (II). Therefore, boric acid at alkaline pH (8) was used as suitable reagent for the precipitation of alizarin red S in aqueous solution and formation of complex between two organic and aqueous solvent. Alizarin has been a well-known ligand for complexation of so many metal ions (Hosseini, 2006; Chamsaz, 2002; Farhadi and Maleki, 2002; Basavaiah, 1999; Amin, 2002; Shama and Amin, 2004; Alkan et al., 2003). So in presence of metal ions, the complex of alizarin red S and boric acid is dissociated and new complex between alizarin red S and metal ions formed and metal solution get red.

Determination of Cu (II) in aqueous solution

Our studies proved that the most sensitive pH for determination Cu (II) by alizarin red S is pH 7.5. So we choose this pH using buffer solution in aliquots of solution of Cu (II) in assay procedure. A linear correlation was found between absorbance and concentration of Cu (II) in the range of 0.6-15 µg.mL⁻¹ (Figure 3). The equation representative calibration curve is:

$$A = 0.0259 C + 0.048$$

$$R^2 = 0.9688$$

Where A and C correspond to absorbance and Cu (II) concentration in µg.mL⁻¹, respectively. The correlation




Figure 3. Calibration curve of Cu (II) at pH 7.5.

Table 1. Effect of foreign ions.

Sn ²⁺	Co ²⁺	Ni ²⁺	Ag ⁺	Hg ₂ ²⁺	Pb ²⁺	metal ions
419	407	404	405	414	410	$\lambda_{\max}$ (nm)
0.0306	0.0291	0.0069	0.0436	0.0385	0.0134	absorbance

coefficient indicates excellent linearity. The molar absorptivity coefficient was found $2.8 \times 10^3 \text{ L.mol}^{-1}.\text{cm}^{-1}$. The method has a detection limit of $0.5 \mu\text{g.mL}^{-1}$.

Effect of foreign ions

The possible interference attributable to various ions, which may be present as the major constituents of mineral samples or found in water samples, was examined by introducing them into the (3 mL) solutions containing $10 \mu\text{g.mL}^{-1}$ of each ion and treating as described in the procedure. The results are given in Table 1. It was observed that all the ions were interfered even at low concentrations. This is due to formation of complex between alizarin red S and the mentioned ions.

Conclusions

The developed method offers a good sensitivity for the determination of Cu (II) in the range of $0.6 - 15 \mu\text{g.mL}^{-1}$ in water. The important feasibility of the method is its simplicity in treatment with the aqueous samples. Results were found not significantly different from AAS and had less standard deviation in this proposed method than in AAS technique.

This paper described a method for the determination of Cu (II) in pure form. In terms of simplicity and expense, the proposed method could be considered superior to the methods reported previously, especially with those based on instrumental analysis. It is clear that alizarin red-S is highly sensitive reagent for the determination of so many metal ions compared with other reagents. Although methods using are more selective compared with the proposed procedure, the latter is less time consuming (maximum time is 10 min), In addition it is applicable at room temperature and has no problem in view of the environ-

mental pollution.

REFERENCES

- Aki MA, Ismael DS, El-Asmy AA (2006). Precipitate flotation-separation,speciation and hydride generation atomic absorption spectrometric determination of selenium(IV) in food stuffs, *Microchem. J.* 83: 61–69.
- Alkan M, Kharun M, Chmilenko F (2003). Spectrophotometric determination of molybdenum with Alizarin Red S in the presence of poly (sulfonylpiperidinylmethyle hydroxide), *Talanta.* 59: 605-611.
- Amin AS (2002). Spectrophotometric determination of piroxicam and tenoxicam in pharmaceutical formulations using alizarin, *J. Pharm. Biomed. Anal.* 29: 729-736.
- Basavaiah K, Latha S, Swamy JM (1999). Spectrophotometric determination of ceterizine hydrochloride with Alizarin Red S, *Talanta.* 50: 887-892.
- Caballero M, Cela R, Perez-Bustamante JA (1990). Analytical applications of some flotation techniques—a review, *Talanta* 37: 275–300.
- Chamsaz M, Hosseini MS, Arbab MH (2002). Synergic flotation spectrophotometric investigation of rare earth(III) ions with Alizarin and 1,10-phenanthroline, *J. Colloid Interface Sci.* 256: 472-476.
- Cundeva K, Stafilov T, Pavlovska G (2000). Flotation separation of cobalt and copper from fresh waters and their determination by electrothermal atomic absorption spectrometry, *Microchem. J.* 65: 165–175.
- Cundeva K, Stafilov T, Pavlovska G (2000). Preconcentration and separation of iron in water by coflotation using lead II hexamethylenedithiocarbamate, *Spectrochim. Acta* 55B: 1081–1087.
- Dadfarnia S, Haji Shabani AM, Jafari AA, Saadat Z, Tamaddon F (2006). Silver, zinc and copper determination in water and biological samples employing FI-FAAS and a microcolumn of immobilized 2,2-[3-aza-1,5-pentanedithiylbis (nitrilomethylidene)]-bisphenol on surfactant coated alumina, *Can. J. Anal. Sci. Spect.* 51: 302–311.
- Duran C, Gundogdu A, Bulut VN, Soylak M, Elci L, Senturk HB, Tufekci M (2007). Solid-phase extraction of Mn(II), Co(II), Ni(II), Cu(II), Cd(II) and Pb(II) ions from environmental samples by flame atomic absorption spectrometry (FAAS), *J. Hazard. Mater.* 146: 347-355.
- Farhadi K, Maleki R (2002). Triiodide ion and alizarin red S as two new reagents for the determination of clotrimazole and ketoconazole, *J. Pharm. Biomed. Anal.* 30: 1023-1033.
- Ghaedi M, Fathi MR, Marahel F, Ahmadi F (2005). Simultaneous preconcentration and determination of copper, nickel, cobalt and lead

- ions content by flame atomic absorption spectrometry, *Fresenius Environ. Bull.* 14: 1158–1163.
- Ghaedi M, Ahmadi F, Karimi H, Gharaghani S (2006). Preconcentration and extraction of copper on activated carbon using 4-amino-2,3-dimethyl-1-phenyl-3-pyrazoline or 4-(4-methoxybenzylidenimin) thiophenole, *J. Korean Chem. Soc.* 50: 23–31.
- Ghaedi M, Asadpour E, Vafaie A (2006). Simultaneous preconcentration and determination of copper, nickel, cobalt, lead, and iron content using a surfactant-coated alumina, *Bull. Chem. Soc. Japan* 79: 432.
- Guo Y, Din B, Lio Y, Chang X, Meng Sh, Tian M (2004). Preconcentration of trace metals with 2-(methylthio) aniline-functionalized XAD-2 and their determination by flame atomic absorption spectrometry, *Anal. Chim. Acta.* 504: 319–324.
- Hosseini MS, Raissi H, Yavari HR (2006). Synergistic flotation of U (VI)–alizarin complex with some diamines followed by spectrophotometric determination of U (VI) using 4, 4'-diaminophenylmethane, *Anal. Chim. Acta.* 559: 181–185.
- Jahromi EZ, Bidari A, Assadi Y, Hosseini MRM, Jamaal MR (2007). Dispersive liquid–liquid microextraction combined with graphite furnace atomic absorption spectrometry: ultra trace determination of cadmium in water samples, *Anal. Chim. Acta* 585: 305–311.
- Karimi H, Ghaedi M, Shokrollahi A, Rajabi HR, Soylak M, Karami B (2008). Development of a selective and sensitive flotation method for determination of trace amounts of cobalt, nickel, copper and iron in environmental samples *Journal of Hazardous Materials* 151: 26–32.
- Kumar M, Rathore DPS, Singh AK (2000). Amberlite XAD-2 functionalized with o-aminophenol: synthesis and applications as extractant for copper(II), cobalt(II), cadmium(II), nickel(II), zinc(II) and lead(II), *Talanta.* 51: 1187–1196.
- Lahiri S, Sarkar S (2007). Separation of iron and cobalt using 59Fe and 60Co by dialysis of polyvinylpyrrolidone–metal complexes: a greener approach, *Appl. Radiat. Isotopes* 65: 385–391.
- Lazaru A, Stafilov T (1993). Determination of Fe, Mn, Cu, Cr and Ni in sulfide minerals from Alshar by atomic absorption spectrometry, *Geol. Macedonica*, 7: 73–80.
- Lazaru A, Stafilov T (1997). Determination of copper in dolomite by electrothermal atomic absorption spectrometry, *Anal. Lab.* 6: 101–103.
- Lemos VA, Baliza PX, Yamaki RT, Rocha ME, Paula A, Alves O (2003). Synthesis and application of a functionalized resin in on-line system for copper preconcentration and determination in foods by flame atomic absorption spectrometry, *Talanta.* 61: 675–682.
- Li QM, Zhao XH, Jiang K, Liu GG (2007). Study of spectrophotometric method for determination of trace copper after the separation and enrichment with solid phase extractant-microcrystalline phenolphthalein, *Chinese Sci. Bull.* 52: 65–70.
- Nakamura T, Okubo K, Sato J (1988). Atomic-absorption spectrometric determination of copper in calcium-carbonate scale and carbonate rocks by direct atomization of solid samples, *Anal. Chim. Acta.* 209: 287–292.
- Paneva VZ, Cundeva K, Stafilov T (2005). Inductively coupled plasma-atomic emission spectrometry method for determination of trace metals in alkaline earth matrices after flotation separation, *Spectrochim. Acta* 60B: 403–407.
- Prabhakaran D, Subramanian MS (2003). A new chelating sorbent for metal ion extraction under high saline conditions, *Talanta.* 59: 1227–1236.
- Rojas FS, Ojeda CB, Pavón JMC (2007). Automated on-line preconcentration of palladium on different sorbents and its determination in environmental samples, *Ann. Chim.* 97: 265–276.
- Safavi A, Abdullah H, Mirzajani R (2006). Determination of Fe(III), Al(III) and Cu(II) by Partial Least-squares Calibration Method, *Spectrochim. Acta. Part A.* 63: 196–199.
- Sakai Y, Tomura T, Ohshita K, Koshimizu S (1998). Determination of trace copper in water samples by neutron activation analysis preceded by preconcentration on activated carbon powder, *J. Radioanal. Nucl. Chem.* 230: 261–263.
- Shama SA, Amin AS (2004). Spectrophotometric microdetermination of nefopam, mebeverine and phenylpropanolamine hydrochloride in pharmaceutical formulations using alizarins, *Spectrochim. Acta. Part A.* 60: 1769–1774.
- Stafilov T (2000). Determination of trace elements in minerals by electrothermal atomic absorption spectrometry, *Spectrochim. Acta* 55: 893–906.
- Stafilov T, Zendelovska D (2000). Determination of cobalt, copper, lead and nickel in gypsum by Zeeman electrothermal atomic absorption spectrometry, *Acta Chim. Slov.* 47: 381–388.
- Stalikas CD (2002). Micelle-mediated extraction as a tool for separation and preconcentration in metal analysis, *Trends Anal. Chem.* 21: 343–355.
- Tewari PK, Singh AK (2001). Synthesis, characterization and applications of pyrocatechol modified Amberlite XAD-2 resin for preconcentration and determination of metal ions in water samples by flame atomic absorption spectrometry (FAAS), *Talanta.* 53: 823–833.
- Venkatesh G, Singh AK (2005). 2-[[1-(3,4 Dihydroxyphenyl) methylidene]amino] benzoic acid immobilized Amberlite XAD-16 as metal extractant, *Talanta.* 67: 187–194.
- Yaman M, Ince M (2006). Solid phase extraction and preconcentration of nickel in mineral waters with PAR-loaded amberlite XAD-7 and determination by flame atomic absorption spectrometry, *Atom. Spectrosc.* 27: 186–192.
- Zhao W, Zhang Y, Li Q (2008). Indirect spectrophotometric determination of sodium ceftriaxone with n-propyl alcohol-ammonium sulfate–water system by extraction flotation of copper(II) *Clinica Chimica Acta* 391: 80–84.
- Zendelovska D, Pavlovska G, Cundeva K, Stafilov T (2001). Electrothermal atomic absorption spectrometric determination of cobalt, copper, lead and nickel traces in aragonite following flotation and extraction separation, *Talanta* 54: 139–146.